

**CERTIFICATE OF COMPLIANCE****LRQA**

hereby declares that the certification assessment has demonstrated that

**GENE-UP® *E. coli* O157:H7**

Unit doses reagents reference 423108 (ECO2 and ECO), kit insert reference 050569, to be used with the software versions 3.0, 3.1, 3.2 and 4.0.

Manufactured by:  
bioMérieux  
Zone Polytec  
5, rue des Berges  
38000 Grenoble  
France

Supplied by:  
bioMérieux SA  
376 Chemin de l'Orme  
69280 Marcy L'Etoile  
France

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report, which has been prepared by a MicroVal Expert Laboratory and approved by the MicroVal Technical Committee. The above-mentioned production facility complies to the QMS requirements as mentioned in the MicroVal Rules. The summary of the validation report is available on the MicroVal website: [www.microval.org](http://www.microval.org).

Reference method: ISO 16654:2001 Microbiology of food and animal feeding stuffs - Horizontal method for the detection of *Escherichia coli* O157

Scope: fresh and frozen raw meats, cured and fermented meats, vegetables, raw dairy products using a specific protocol and raw meats (except poultry) using a short protocol (25 g and 375 g sample size), and environmental samples

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules.

Certificate no.: 2015LR59

First approval date: 06 April 2017

Surveillance date: 22 July 2025

Expiry date: 05 April 2029



---

ISSUED BY:

LRQA Nederland B.V.  
Rotterdam, The Netherlands