

**CERTIFICATE OF COMPLIANCE****LRQA**

hereby declares that the certification assessment has demonstrated that

**COMPACT DRY ETC**

for the detection and enumeration of *Enterococcus*

Manufactured and supplied by:  
Shimadzu Diagnostics Corporation  
20th Floor Ueno Frontier Tower  
3-24-6 Ueno, Taito-ku, Tokyo, 110-8736  
JAPAN

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report, which has been prepared by a MicroVal Expert Laboratory and approved by the MicroVal Technical Committee. The above-mentioned production facility complies to the QMS requirements as mentioned in the MicroVal Rules. The summary of the validation report is available on the MicroVal website: [www.microval.org](http://www.microval.org).

Reference method: NMKL Method No. 68 5th Edition 2011: *Enterococcus*. Determination in foods and feeds

Scope: a broad range of foods

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules.

Certificate no.: 2014LR48

First approval date: 10 April 2017

Surveillance date: 19 June 2025

Expiry date: 31 December 2025



---

ISSUED BY: LRQA Nederland B.V.  
Rotterdam, The Netherlands