



CERTIFICATE OF COMPLIANCE

LRQA

hereby declares that the certification assessment has demonstrated that

MC-Media Pad™ ACplus

Manufactured and supplied by:

JNC Corporation

2-2-1 Otemachi

Chiyoda-ku

Tokyo

Japan, 100-8105

Production site:

JNC Corporation

Yokohama Research Center

5-1, Ookawa

Kanazawa-ku

Yokohama

Japan, 236-8605

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website: www.microval.org

Reference method: ISO 4833-1:2013 Microbiology of the food chain —

Horizontal method for the enumeration of microorganisms Part 1: Colony count at 30 degrees C by the pour plate technique

Scope: A broad range of foods

The validation and certification has been performed in accordance with EN ISO 16140-2: 2016, EN ISO 16140-2:2016/Amd. 1:2024 and the MicroVal Rules.

This certificate is valid only in association with the certificate schedule bearing the same number on which the distributor applicable to this approval is listed.

Certificate no.: 2015LR52

First approval date: 29 January 2018

Renewal date: 22 July 2026

Expiry date: 28 January 2030

ISSUED BY: LRQA Nederland B.V.
Rotterdam, The Netherlands



CERTIFICATE SCHEDULE

MC-Media Pad™ ACplus

Manufactured and supplied by:

JNC Corporation
2-2-1 Otemachi
Chiyoda-ku
Tokyo
Japan, 100-8105

Production site:

JNC Corporation
Yokohama Research Center
5-1, Ookawa
Kanazawa-ku
Yokohama
Japan, 236-8605

MC-Media Pad™ ACplus is also supplied as MC-Media Pad™ Rapid Aerobic Count by:

Merck KGaA
Frankfurter Straße 250
64293 Darmstadt
Germany

Certificate no.: 2015LR52

First approval date: 29 January 2018

Renewal date: 22 July 2026

Expiry date: 28 January 2030