

CERTIFICATE OF COMPLIANCE**LRQA**

hereby declares that the certification assessment has demonstrated that

Fossomatic™ 7DC

for enumeration of *somatic cells* in raw cow's milk

Manufactured and supplied by:

FOSS ANALYTICAL A/S
FOSS ALLÉ
DK-3400 HILLERØD
DENMARK

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report, which has been prepared by a MicroVal Expert Laboratory and approved by the MicroVal Technical Committee. The above-mentioned production facility complies to the QMS requirements as mentioned in the MicroVal Rules. The summary of the validation report is available on the MicroVal website: www.microval.org.

Reference method: ISO 13366-1:2008 Milk - Enumeration of *somatic cells* - Part 1:
Microscopic method (Reference method)

Scope: raw cow's milk

The validation and certification has been performed in accordance with ISO 8196-3, ISO 13366-2, the EURL MMP criteria - Criteria for the validation of instrumental (epifluorescent) methods for the enumeration of somatic cells in raw cow's milk, version 2, 21/01/2013 and the MicroVal Rules.

Certificate no.: 2016LR63

First approval date: 12 June 2018

Renewal date: 09 January 2026

Expiry date: 11 June 2030



ISSUED BY: LRQA Nederland B.V.
Rotterdam, The Netherlands