



CERTIFICATE OF COMPLIANCE

LRQA

hereby declares that the certification assessment has demonstrated that

MALDI Biotyper®

MBT microflex LT/SH / MBT smart/ MBT sirius one / MBT sirius
Complete Solution for the confirmation of *Cronobacter* spp.

Manufactured and supplied:
Bruker Daltonics GmbH & Co. KG
Fahrenheitstraße 4
28359 Bremen
Germany

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website: www.microval.org

Reference method:

- ISO 22964 (2017) - Microbiology of the food chain - Horizontal method for the detection of *Cronobacter* spp.

Scope: Confirmation of *Cronobacter* spp. from colonies isolated on ESIA, CCI and any non-selective nutrient agars

The validation and certification has been performed in accordance with ISO 16140-6:2019 and the MicroVal Rules.

Certificate no.: 2017LR72

First approval date: 1 March 2018

Renewal date: 20 January 2026

Expiry date: 28 February 2030

ISSUED BY:

LRQA Nederland B.V.
Rotterdam, The Netherlands



Year modification	Description modification	Date approval MVTC
2026	Renewal of certificate, database 1829023 "MBT Compass Library 2023"	January 2026