



CERTIFICATE OF COMPLIANCE

LRQA

hereby declares that the certification assessment has demonstrated that

Kylt® Salmonella spp. 2.0

Manufactured and supplied by:
SAN Group Biotech Germany GmbH
Muehlenstraße 13
49685 Hoeltinghausen
Germany

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report, which has been prepared by a MicroVal Expert Laboratory and approved by the MicroVal Technical Committee. The above-mentioned production facility complies to the QMS requirements as mentioned in the MicroVal Rules. The summary of the validation report is available on the MicroVal website: www.microval.org.

Reference method: ISO 6579-1:2017 Microbiology of the food chain — Horizontal method for the detection, enumeration and serotyping of *Salmonella* - Part 1: Detection of *Salmonella* spp.

Scope: raw meat and ready to cook meat products, raw poultry and ready to cook poultry products, environmental samples, primary production samples

The validation and certification has been performed in accordance with EN ISO 16140-2:2016, EN ISO 16140-2:2016/Amd. 1:2024 and the MicroVal Rules.

Certificate no.: 2017LR78

First approval date: 06 July 2021
Surveillance date: 17 February 2026
Expiry date: 05 July 2029

ISSUED BY: _____
LRQA Nederland B.V.
Rotterdam, The Netherlands

| DNA extraction PCR profile combination | Category | Platforms certified | | | | | | |
|----------------------------------------|---------------------------------|---------------------|-----|-----|-----|-----|-----|-----|
| | | A | B | C | D | E | F | G |
| Extraction mix II profile I | Meat 25g | Yes | Yes | Yes | Yes | Yes | Yes | Yes |
| | Meat 375g | Yes | Yes | Yes | Yes | Yes | Yes | Yes |
| | Poultry 25g | Yes | Yes | Yes | Yes | Yes | Yes | Yes |
| | Poultry 375g | Yes | Yes | Yes | Yes | Yes | Yes | Yes |
| | Environmental | Yes | Yes | Yes | Yes | Yes | Yes | Yes |
| | Primary production | Yes | Yes | Yes | Yes | Yes | Yes | Yes |
| | Primary production 5h RVS broth | Yes | Yes | Yes | Yes | Yes | Yes | Yes |

Key A =7500 FAST B = CFX96 C = CFX 384, D = Rotor Gene 3000, E = Rotor Gene 6000, F= Aria MX Pro, G= Light cycler 480

| DNA extraction PCR profile combination | Category | Platforms certified | | | | | | |
|----------------------------------------|---------------------------------|---------------------|-----|-----|-----|-----|-----|-----|
| | | A | B | C | D | E | F | G |
| Extraction mix II profile II | Meat 25g | yes | yes | yes | yes | yes | yes | yes |
| | Meat 375g | yes | yes | yes | yes | yes | yes | yes |
| | Poultry 25g | yes | yes | yes | yes | yes | yes | yes |
| | Poultry 375g | yes | yes | yes | yes | yes | yes | yes |
| | Environmental | yes | yes | yes | yes | yes | yes | yes |
| | Primary production | yes | yes | yes | yes | yes | yes | yes |
| | Primary production 5h RVS broth | yes | yes | yes | yes | yes | yes | yes |

Key A =7500 FAST B = CFX96 C = CFX 384, D = Rotor Gene 3000, E = Rotor Gene 6000, F= Aria MX Pro, G= Light cycler 480

| DNA extraction PCR profile combination | Category | Platforms certified | | | | | | |
|----------------------------------------|---------------------------------|---------------------|-----|-----|-----|-----|-----|-----|
| | | A | B | C | D | E | F | G |
| RNA/DNA extraction profile I | Meat 25g | yes | yes | no | yes | yes | yes | no |
| | Meat 375g | yes | yes | no | yes | yes | yes | no |
| | Poultry 25g | yes | yes | yes | yes | yes | yes | yes |
| | Poultry 375g | yes | yes | yes | yes | yes | yes | yes |
| | Environmental | yes | yes | yes | yes | yes | yes | yes |
| | Primary production | yes | yes | yes | yes | yes | yes | yes |
| | Primary production 5h RVS broth | yes | yes | yes | yes | yes | yes | yes |

Key A =7500 FAST B = CFX96 C = CFX 384, D = Rotor Gene 3000, E = Rotor Gene 6000, F= Aria MX Pro, G= Light cycler 480

| DNA extraction PCR profile combination | Category | Platforms certified | | | | | | |
|----------------------------------------|---------------------------------|---------------------|-----|-----|-----|-----|-----|-----|
| | | A | B | C | D | E | F | G |
| RNA/DNA extraction profile II | Meat 25g | yes | yes | yes | no | yes | yes | no |
| | Extraction mix II profile I | yes | yes | yes | no | yes | yes | no |
| | Poultry 25g | yes | yes | yes | yes | yes | yes | yes |
| | Poultry 375g | yes | yes | yes | Yes | yes | yes | yes |
| | Environmental | yes | yes | yes | Yes | yes | yes | yes |
| | Primary production | yes | yes | yes | Yes | yes | yes | yes |
| | Primary production 5h RVS broth | yes | yes | yes | yes | yes | yes | yes |

Key A =7500 FAST B = CFX96 C = CFX 384, D = Rotor Gene 3000, E = Rotor Gene 6000, F= Aria MX Pro, G= Light cycler 480