

CERTIFICATE OF COMPLIANCE

LRQA

hereby declares that the certification assessment has demonstrated that

iQ-Check STEC VirX Kit

Corporate location:
Bio-Rad Laboratories
2000 Nobel Drive
Hercules, CA 94547
USA

Manufactured by:
Bio-Rad Laboratories
925 Nobel Drive
Hercules, CA 94547
USA

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report, which has been prepared by a MicroVal Expert Laboratory and approved by the MicroVal Technical Committee. The above-mentioned production facility complies to the QMS requirements as mentioned in the MicroVal Rules. The summary of the validation report is available on the MicroVal website:

www.microval.org.

Reference method: ISO/TS 13136:2012 Microbiology of food and animal feed - Real-Time Polymerase Chain Reaction (PCR) based method for the detection of food-borne pathogens. Horizontal method for the detection of shiga-toxin producing (STEC) and the determination of O26, O103, O111, O145 and O157 serogroups

Scope: a broad range of foods including raw dairy products (up to 25 g test portion), raw meat products (excluding poultry) (up to 375g test portion), flour & raw doughs (up to 375 test portion), multicomponent foods and meal components (up to 25 g test portion), raw produces and fruits (up to 25 g test portion)

The validation and certification has been performed in accordance with EN ISO 16140-2:2016, EN ISO 16140-2:2016/Amd. 1:2024 and the MicroVal Rules.

Certificate no.: 2021LR96

First approval date: 27 June 2022

Surveillance date: 09 June 2026

Expiry date: 26 June 2030



ISSUED BY:

LRQA Nederland B.V.
Rotterdam, The Netherlands

Year	Description	Date approval MVTC
2026	Renewal of the certificate, IFU 10000131516 Ver D October 2024	69 th MVTC meeting, 09 June 2026