

CERTIFICATE OF COMPLIANCE

LRQA

hereby declares that the certification assessment has demonstrated that

Check & Trace Salmonella 2.0 (CTS 2.0)

Manufactured and supplied by:
Check-Points B.V.
Binnenhaven 5
6709 PD Wageningen
The Netherlands

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website:

www.microval.org

Reference methods:

- ISO6579-1:2017 – Microbiology of the food chain – Horizontal method for the detection, enumeration and serotyping of *Salmonella* – Part 1: Detection of *Salmonella* spp.
- ISO/TR 6579-3:2014 - Microbiology of the food chain – Horizontal method for the detection, enumeration and serotyping of *Salmonella* – Part 3: Guidelines for serotyping *Salmonella* spp.
- ISO6579-4:2025 Microbiology of the food chain - Horizontal method for the detection, enumeration and serotyping of *Salmonella* - Part 4: Identification of monophasic *Salmonella* Typhimurium (1,4,[5],12:i:-) by polymerase chain reaction (PCR).

Scope: Confirmation of presumptive *Salmonella* isolated on non selective NA and selective XLD agar and typing of 106 *Salmonella* serovars.

The validation and certification has been performed in accordance with EN ISO 16140-6:2019 and the MicroVal Rules.

This certificate is only valid in conjunction with attachment Conclusions MCS ILS CTS 2.0 v8.

Certificate no.: 2021LR107

First approval date: 2 December 2022

Surveillance date: 17 September 2025

Expiry date: 1 December 2026



ISSUED BY: LRQA Nederland B.V.
Rotterdam, The Netherlands

Year modification	Description modification	Date approval MVTC
2025	Extension of the scope, 59 to 106 <i>Salmonella</i> serovars	17 September 2025