

**CERTIFICATE OF COMPLIANCE****LRQA**

hereby declares that the certification assessment has demonstrated that

**D-COUNT® Commercial Sterility Testing**

IFUs referenced DC25\_50-D720C-01 and DC25\_50-D714C-01

Manufactured by:

bioMérieux SA

4 bis Rue des Coutures

36270 Combourg

France

Supplied by:

bioMérieux SA

376 Chemin de l'Orme

69280 Marcy L'Etoile

France

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report, which has been prepared by a MicroVal Expert Laboratory and approved by the MicroVal Technical Committee. The above-mentioned production facility complies to the QMS requirements as mentioned in the MicroVal Rules. The summary of the validation report is available on the MicroVal website: [www.microval.org](http://www.microval.org)


Reference method: Codex Code of Hygienic practice for milk and milk products (CAC/RCP 57-2004). Appendix B: Microbiocidal control measures. Section 2.2: Process management for UHT treatment using ISO4833-1:2013 for the plate count at 30°C.

Scope: UHT treated milks and plant-based drinks up to 1 L. (UHT liquid dairy products, UHT plant-based drinks and UHT flavoured milks)

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules.

Certificate no.: 2023LR126

First approval date: 19 February 2025  
Expiry date: 18 February 2029



---

ISSUED BY: LRQA Nederland B.V.  
Rotterdam, The Netherlands