

CERTIFICATE OF COMPLIANCE**LRQA**

hereby declares that the certification assessment has demonstrated that

BactoScan™ 5

Manufactured and supplied by:

Foss Analytical A/S

Foss Allé

DK-3400 Hillerød

Denmark

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report, which has been prepared by a MicroVal Expert Laboratory and approved by the MicroVal Technical Committee. The above-mentioned production facility complies to the QMS requirements as mentioned in the MicroVal Rules. The summary of the validation report is available on the MicroVal website: www.microval.org

Reference method: EN-ISO 4833-1:2013 Microbiology of the food chain – Horizontal method for the enumeration of microorganisms – Part 1: Colony count at 30 degrees C by the pour plate technique


Scope: raw cow milk, raw sheep and raw goat milk

The validation and certification has been performed in accordance with EN ISO 16140-2:2016, ISO 16297:2020, ISO 21187:2021 and the EURL for Milk and Milk Products document (Criteria for the validation of instrumental (epifluorescent) methods for the determination of total flora in raw milk – Version 2 – [21 December 2011]) and the MicroVal Rules.

Certificate no.: 2024LR133

First approval date: 10 January 2025

Expiry date: 9 January 2029



ISSUED BY: LRQA Nederland B.V.
Rotterdam, The Netherlands

| Year modification | Description modification | Date approval MVTC |
|-------------------|---|--------------------|
| 2025 | Scope was extended with raw sheep and goat milk | 20 June 2025 |