

CERTIFICATE OF COMPLIANCE**LRQA**

hereby declares that the certification assessment has demonstrated that

Autof ms1000

Manufactured and supplied by:
Autobio Diagnostics Co., Ltd
No. 199, 15th Ave
National Eco & Tech Development Area
Zhengzhou
China

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report, which has been prepared by a MicroVal Expert Laboratory and approved by the MicroVal Technical Committee. The above-mentioned production facility complies to the QMS requirements as mentioned in the MicroVal Rules. The summary of the validation report is available on the MicroVal website: www.microval.org

Reference methods:

- ISO 6579-1:2017 Microbiology of the food chain – Horizontal method for detection, enumeration and serotyping of *Salmonella* – Part 1: Detection of *Salmonella* spp.;
- ISO 6579-1:2017/Amd 1:2020 – Microbiology of the food chain – Horizontal method for the detection, enumeration and serotyping of *Salmonella* – Part 1: Detection of *Salmonella* spp.


Scope: the confirmation of presumptive *Salmonella* spp. colonies isolated from selective and non-selective agars and under specified conditions for a genus level claim.

The validation and certification has been performed in accordance with EN ISO 16140-6:2019 and the MicroVal Rules.

Certificate no.: 2023LR116

First approval date: 1 May 2025

Expiry date: 30 April 2029



ISSUED BY: LRQA Nederland B.V.
Rotterdam, The Netherlands