



## CERTIFICATE OF COMPLIANCE

### LRQA

hereby declares that the certification assessment has demonstrated that

### **Salmofast® PCR method**

Manufactured and supplied by:  
Microbial Systems S.L.  
C/Pic de Peguera 15  
Edifici Jaume Casademont E  
17003 Girona  
Spain

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report, which has been prepared by a MicroVal Expert Laboratory and approved by the MicroVal Technical Committee. The above-mentioned production facility complies to the QMS requirements as mentioned in the MicroVal Rules. The summary of the validation report is available on the MicroVal website: [www.microval.org](http://www.microval.org)

#### Reference method:

1. ISO 6579-1:2017 Microbiology of the food chain - Horizontal method for the detection, enumeration and serotyping of *Salmonella* - Part 1: Detection of *Salmonella* spp. *Annex D was not carried out during the validation study.*
2. ISO 6579-1/A1:2020 Microbiology of the food chain - Horizontal method for the detection, enumeration and serotyping of *Salmonella* - Part 1: Detection of *Salmonella* spp. - Amendment 1 Broader range of incubation temperatures, amendment to the status of Annex D, and correction of the composition of MSRV and SC.

Scope: broad range of food products (excluding raw poultry) – up to 25g, production environmental samples (excluding dust) – up to 25g or ml or sampling devices

The validation and certification has been performed in accordance with EN ISO 16140-2:2016, EN ISO 16140-2:2016/Amd. 1:2024 and the MicroVal Rules.

Certificate no.: 2023LR125

First approval date: 06 May 2026

Expiry date: 05 May 2030

ISSUED BY:

\_\_\_\_\_  
LRQA Nederland B.V.  
Rotterdam, The Netherlands



Year modification	Description modification	Date approval MVTC
2026	IFU Salmofast Rev. 20, 01/2026	67 <sup>th</sup> MVTC meeting